


Уплуші HealthCare Device

A product range for your perfect wellness measuring.



Pulse Oximeter





Series: SXY • Model No. : FLBK06

Technical Specification			
SpO2		Power Requirements	
Measure range	35-99%	Battery	2 Batteries with 1.5 volts of AAA sizes
Accuracy	±1% in the range of 90%-99%		2 x 1.5v AAA
	±2% in the range of 70%-89%	Environment Requirements	
Pulse Rate		Operation temperature	5-~0 °C
Measure range	30 bpm- 240 bpm	Storage temperature	-10~40 °C
Resolution	1 bpm	Ambient humadity	15-80%,
Accuracy	±2 bpm		No condition in operation 10%-93%
Pulse intensity	Bargraph		No condensation in storage



We support make in India movement

Manufactured by: รัสกรูป Electronics Pvt. Ltd

Plot No.- E-59, Market Yard, Pune, Maharashtra, India

www.sansuiscales.com

Mob: +91-9823-888-888 / 7767900040 | Email: sales@sansui.co.in / Support@Sansui.co.in

